

Contents

1. Description
 - 1.1 Principle of the MACS® Separation
 - 1.2 Background information
 - 1.3 Applications
 - 1.4 Reagent and instrument requirements
2. Protocol
 - 2.1 Sample preparation
 - 2.2 Magnetic labeling
 - 2.3 Magnetic separation
 - 2.4 Cell separation with the autoMACS® Pro Separator
3. Example of a separation using CD45 (TIL) MicroBeads

Warnings

Reagents contain sodium azide. Under acidic conditions sodium azide yields hydrazoic acid, which is extremely toxic. Azide compounds should be diluted with running water before discarding. These precautions are recommended to avoid deposits in plumbing where explosive conditions may develop.

1. Description

This product is for research use only.

Components	1 mL CD45 (TIL) MicroBeads, mouse: MicroBeads conjugated to monoclonal anti-mouse CD45 antibodies (isotype: rat IgG2b).
Capacity	For 10 ⁹ total cells, up to 100 separations.
Product format	CD45 (TIL) MicroBeads are supplied in buffer containing stabilizer and 0.05% sodium azide.
Storage	Store protected from light at 2–8 °C. Do not freeze. The expiration date is indicated on the vial label.

1.1 Principle of the MACS® Separation

First, the CD45⁺ cells are magnetically labeled with CD45 (TIL) MicroBeads. Then, the cell suspension is loaded onto a MACS® Column, which is placed in the magnetic field of a MACS Separator. The magnetically labeled CD45⁺ cells are retained within the column. The unlabeled cells run through; this cell fraction is thus depleted of CD45⁺ cells. After removing the column from the magnetic field, the magnetically retained CD45⁺ cells can be eluted as the positively selected cell fraction. To increase the purity, the positively selected cell fraction containing the CD45⁺ cells must be separated over a second column.

1.2 Background information

CD45 (TIL) MicroBeads, mouse have been developed for the isolation of tumor-infiltrating leukocytes (TILs) from single-cell suspensions of solid mouse tumors. The CD45 antigen is expressed on all cells of hematopoietic origin except erythrocytes and platelets.

1.3 Applications

- Positive selection of CD45⁺ leukocytes from solid mouse tumors, e.g., B16F10, 4T1, or CT26.WT.

1.4 Reagent and instrument requirements

- Buffer: Prepare a solution containing phosphate-buffered saline (PBS), pH 7.2, 0.5% bovine serum albumin (BSA), and 2 mM EDTA by diluting MACS® BSA Stock Solution (# 130-091-376) 1:20 with autoMACS® Rinsing Solution (# 130-091-222). Keep buffer cold (2–8 °C). Degas buffer before use, as air bubbles could block the column. Always use freshly prepared buffer.

▲ **Note:** EDTA can be replaced by other supplements such as anticoagulant citrate dextrose formula-A (ACD-A) or citrate phosphate dextrose (CPD). BSA can be replaced by other proteins such as mouse serum albumin, mouse serum, or fetal bovine serum (FBS). Buffers or media containing Ca²⁺ or Mg²⁺ are not recommended for use.
- MACS Columns and MACS Separators: CD45⁺ cells can be enriched by using MS or LS Columns. Positive selection can also be performed by using the autoMACS Pro or the MultiMACS™ Cell24 Separator.

Column	Max. number of labeled cells	Max. number of total cells	Separator
Positive selection			
MS	10 ⁷	2×10 ⁷	MiniMACS, OctoMACS, VarioMACS, SuperMACS II
LS	4×10 ⁷	5×10 ⁷	MidiMACS, QuadroMACS, VarioMACS, SuperMACS II
	4×10 ⁷	5×10 ⁷	MultiMACS Cell24 Separator Plus
autoMACS	5×10 ⁷	10 ⁸	autoMACS Pro
Multi-24 Column Block (per column)	2×10 ⁷	2.5×10 ⁷	MultiMACS Cell24 Separator Plus

▲ **Note:** Column adapters are required to insert certain columns into the VarioMACS™ or SuperMACS™ II Separators. For details refer to the respective MACS Separator data sheet.

▲ **Note:** If separating with LS Columns and the MultiMACS Cell24 Separator Plus use the Single-Column Adapter. Refer to the user manual for details.

- Tumor Dissociation Kit, mouse (# 130-096-730) for the generation of single-cell suspension from tumor tissues.

- gentleMACS™ Dissociator (# 130-093-235), gentleMACS Octo Dissociator (# 130-095-937), or gentleMACS Octo Dissociator with Heaters (# 130-096-427)
- gentleMACS C Tubes (# 130-093-237, # 130-096-334)
- (Optional) Fluorochrome-conjugated REA (REAffinity™ antibodies: recombinantly engineered, lacking Fcγ-binding site) CD45 antibodies for flow cytometric analysis, e.g., CD45-VioBlue®. For more information about antibodies refer to www.miltenyibiotec.com/antibodies.
▲ **Note:** Due to expression of Fcγ receptors on tumor-infiltrating leukocytes REA antibodies are recommended.
- (Optional) Propidium Iodide Solution (# 130-093-233), DAPI Staining Solution (# 130-111-570), 7-AAD Staining Solution (# 130-111-568), or Viability™ Fixable Dyes (# 130-109-812, # 130-109-814, # 130-109-816) for flow cytometric exclusion of dead cells.
- (Optional) Dead Cell Removal Kit (# 130-090-101) for the depletion of dead cells.
- (Optional) Pre-Separation Filters (30 μm) (# 130-041-407) to remove cell clumps.
- (Optional) MACS SmartStrainers (30 μm) (# 130-098-458) to remove cell clumps.

2. Protocol

2.1 Sample preparation

For preparation of a single-cell suspension from solid mouse tumors use the Tumor Dissociation Kit, mouse (# 130-096-730) in combination with the gentleMACS™ Dissociators.

For details refer to www.gentlemacs.com/protocols.

▲ Dead cells may bind non-specifically to MACS® MicroBeads. To remove dead cells, we recommend using the Dead Cell Removal Kit (# 130-090-101).



2.2 Magnetic labeling

▲ Cells can be labeled with MACS MicroBeads using the autolabeling function of the autoMACS® Pro Separator. For more information refer to section 2.4.

▲ Work fast, keep cells cold, and use pre-cooled solutions. This will prevent capping of antibodies on the cell surface and non-specific cell labeling.

▲ Volumes for magnetic labeling given below are for up to 10^7 total cells. When working with fewer than 10^7 cells, use the same volumes as indicated. When working with higher cell numbers, scale up all reagent volumes and total volumes accordingly (e.g. for 2×10^7 total cells, use twice the volume of all indicated reagent volumes and total volumes).

▲ For optimal performance it is important to obtain a single-cell suspension before magnetic labeling. Pass cells through 30 μm nylon mesh (MACS SmartStrainers (30 μm), # 130-098-458). Moisten filter with buffer before use.

▲ The recommended incubation temperature is 2–8 °C. Higher temperatures and/or longer incubation times may lead to non-specific cell labeling. Working on ice may require increased incubation times.

1. Determine cell number.
2. Centrifuge cell suspension at 300×g for 5 minutes. Aspirate supernatant completely.
3. Resuspend cell pellet in 90 μL of buffer per 10^7 total cells.
▲ **Note:** Always use freshly prepared buffer.
4. Add 10 μL of CD45 (TIL) MicroBeads per 10^7 total cells.
5. Mix well and incubate for 15 minutes in the dark in the refrigerator (2–8 °C).
6. (Optional) Add staining antibodies according to manufacturer's recommendations.
7. Add buffer to a final volume of 500 μL for up to 5×10^7 cells.
▲ **Note:** If more cells were used, split the sample onto multiple columns during magnetic separation.
▲ **Note:** For higher cell numbers, scale up buffer volume accordingly.
8. Proceed to magnetic separation (2.3).



2.3 Magnetic separation

▲ Choose an appropriate MACS Column and MACS Separator according to the number of total cells and the number of CD45⁺ cells. For details refer to the table in section 1.4.

▲ **Note:** MS Columns are recommended for highest purity of CD45⁺ cells. LS Columns are recommended for highest recovery of CD45⁺ cells.

▲ For optimal performance it is important to obtain a single-cell suspension before magnetic separation. Pass cells through 30 μm nylon mesh (Pre-Separation Filters (30 μm), # 130-041-407) to remove cell clumps which may clog the column. Moisten filter with buffer before use.

▲ Always wait until the column reservoir is empty before proceeding to the next step.

Magnetic separation with MS or LS Columns

1. Place column in the magnetic field of a suitable MACS Separator. For details refer to the respective MACS Column data sheet.
2. Prepare column by rinsing with the appropriate amount of buffer:
MS: 500 μL LS: 3 mL
3. Apply cell suspension onto the column. Collect flow-through containing unlabeled cells.
4. Wash column with the appropriate amount of buffer. Collect unlabeled cells that pass through and combine with the flow-through from step 3.
MS: 3×500 μL LS: 2×1 mL
▲ **Note:** Perform washing steps by adding buffer aliquots as soon as the column reservoir is empty.
5. Remove column from the separator and place it on a suitable collection tube.

- Pipette the appropriate amount of buffer onto the column. Immediately flush out the magnetically labeled cells by firmly pushing the plunger into the column.

MS: 1 mL LS: 3 mL

- (Optional) To increase the purity of CD45⁺ cells, the eluted fraction can be enriched over a second MS or LS Column. Repeat the magnetic separation procedure as described in steps 1 to 6 by using a new column.

Magnetic separation with the MultiMACS™ Cell24 Separator

Refer to the the MultiMACS™ Cell Separator user manual for instructions on how to use the MultiMACS Cell24 Separator.

2.4 Cell separation with the autoMACS® Pro Separator

▲ Refer to the user manual for instructions on how to use the autoMACS® Pro Separator.

▲ All buffer temperatures should be $\geq 10^{\circ}\text{C}$.

▲ For appropriate resuspension volumes and cell concentrations, please visit www.automacspro.com/autolabeling.

▲ Place tubes in the following Chill Rack positions:

position A = sample, position B = negative fraction,

position C = positive fraction.

2.4.1 Fully automated cell labeling and separation

- Switch on the instrument for automatic initialization.
- Go to the **Reagent** menu and select **Read Reagent**. Scan the 2D barcode of each reagent vial with the barcode scanner on the autoMACS® Pro Separator. Place the reagent into the appropriate position on the reagent rack.
- Place sample and collection tubes into the Chill Rack.
- Go to the **Separation** menu and select the reagent name for each sample from the **Labeling** submenu (the correct labeling, separation, and wash protocols will be selected automatically).
- Enter sample volume into the **Volume** submenu. Press **Enter**.
- Select **Run**.


2.4.2 Magnetic separation using manual labeling

- Label the sample as described in section 2.2 Magnetic labeling.
- Prepare and prime the instrument.
- Apply tube containing the sample and provide tubes for collecting the labeled and unlabeled cell fractions. Place sample and collection tubes into the Chill Rack.
- For a standard separation choose one the following programs:
Positive selection:
Posseld2 for highest purity
or
Possels for highest recovery
Collect positive fraction in row C of the tube rack.

3. Example of a separation using CD45 (TIL) MicroBeads

A tumor induced by the B16F10 cell line was dissociated using the gentleMACS™ Octo Dissociator with Heaters in combination with the Tumor Dissociation Kit, mouse. CD45⁺ TILs were isolated from the single-cell suspension using CD45 (TIL) MicroBeads, an MS Column, and a MiniMACS™ Separator.

Cells were fluorescently stained with CD45-PE and Labeling-Check-Reagent-VioBlue® and analyzed by flow cytometry using the MACSQuant® Analyzer. Cell debris and dead cells were excluded from the analysis based on scatter signals and propidium iodide fluorescence.



Refer to www.miltenyibiotec.com for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit www.miltenyibiotec.com/local to find your nearest Miltenyi Biotec contact.

Legal notices

Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at www.miltenyibiotec.com, as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit www.miltenyibiotec.com for the most up-to-date information on Miltenyi Biotec products.

Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at www.miltenyibiotec.com for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

Trademarks

autoMACS, gentleMACS, MACS, MACSQuant, MidiMACS, the Miltenyi Biotec logo, MiniMACS, MultiMACS, OctoMACS, QuadroMACS, REAfinity, SuperMACS, VarioMACS, Viobility, and VioBlue are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide.

Copyright © 2020 Miltenyi Biotec and/or its affiliates. All rights reserved.