


Index

1. Description
 - 1.1 Principle of Anti-PE MultiSort Kit
 - 1.2 Background and product applications
 - 1.3 Reagent and instrument requirements
2. Protocol
 - 2.1 Sample preparation
 - 2.2 Magnetic labeling
 - 2.3 Magnetic separation
 - 2.4 Removal of MultiSort MicroBeads using MultiSort Release Reagent
 - 2.5 Recommendations to optimize cell isolation using Anti-PE MultiSort Kit
3. Example of a separation using Anti-PE MultiSort Kit

1. Description

Components	2 mL Anti-PE MultiSort MicroBeads: MultiSort MicroBeads conjugated to monoclonal mouse anti-PE (phycoerythrin) antibodies (isotype: mouse IgG1; clone: PE4-14D10) 1 mL MultiSort Release Reagent 2 mL MultiSort Stop Reagent
Size	For 1–5×10 ⁹ total cells
Product format	The Anti-PE MultiSort MicroBeads are supplied as a suspension containing stabilizer and 0.05% sodium azide. The MultiSort Release Reagent is supplied in a solution containing stabilizer and 0.05% sodium azide. The MultiSort Stop Reagent is supplied in a solution containing 0.05% sodium azide.
Storage	Store protected from light at 4–8 °C. Do not freeze. The expiration date is indicated on the vial label.

1.1 Principle of Anti-PE MultiSort Kit



- 1 Cells are labeled with PE-conjugated primary antibodies.
- 2 Cells are magnetically labeled with Anti-PE MultiSort MicroBeads.
- 3 Labeled cells are enriched by using MACS® Columns and Separators.
- 4 The MultiSort Release Reagent releases the MultiSort MicroBeads from the cells.
- 5 Remaining magnetically labeled cells are removed by using MACS Columns and Separators. The released cells are collected in the flow through.
- 6 The release reaction is stopped by adding MultiSort Stop Reagent to the flow-through fraction.
- 7 Magnetic labeling of a second cell surface marker is performed by using MACS MicroBeads.
- 8 Labeled cells are enriched by using MACS Columns and Separators.

1.2 Background and product applications

The Anti-PE MultiSort Kit is an indirect magnetic labeling system which allows the sorting of cells according to multiple surface markers. Cells are first stained with a PE-conjugated primary antibody followed by magnetic labeling with Anti-PE MultiSort MicroBeads. Following enrichment of the PE-labeled cells, using a MACS Column and Separator, the magnetic particles are removed from the cells by using MultiSort Release Reagent. This allows a second magnetic labeling and separation of the cells for another surface marker of interest. Magnetic labeling for the second marker is achieved by using either direct or indirect magnetic labeling with MACS® MicroBeads. For indirect labeling, Anti-Biotin MicroBeads, Anti-FITC MicroBeads or Anti-APC MicroBeads are recommended.

1.3 Reagent and instrument requirements

- Buffer (degassed): PBS (phosphate buffered saline) pH 7.2, supplemented with 0.5% BSA (bovine serum albumin) and 2 mM EDTA. Keep buffer cold (4–8 °C).

▲ **Note:** EDTA can be replaced by other supplements such as anticoagulant citrate dextrose formula-A (ACD-A) or citrate phosphate dextrose (CPD). BSA can be replaced by other proteins such as gelatine, human serum albumin, human serum or fetal calf serum. Buffers or media containing Ca^{2+} or Mg^{2+} are not recommended for use.

- PE-conjugated antibody, peptide or ligand.

▲ **Note:** In particular cases PE-conjugated antibodies from other suppliers than Miltenyi Biotec may not be suitable.

- MACS Columns and MACS Separators:

Table 1: MACS Columns and MACS Separators

Column	max. number of labeled cells	max. number of total cells	Separator
MS	10^7	2×10^8	MiniMACS, OctoMACS, VarioMACS, SuperMACS
LS	10^8	2×10^9	MidiMACS, QuadroMACS, VarioMACS, SuperMACS
autoMACS	2×10^8	4×10^9	autoMACS

▲ **Note:** Column adapters are required to insert MS or LS Columns into VarioMACS or SuperMACS. For details, see MACS Separator data sheets.

- (Optional) Pre-Separation Filters (# 130-041-407) to remove cell clumps.

2. Protocol

2.1 Sample preparation

When working with anticoagulated peripheral blood or buffy coat, peripheral blood mononuclear cells (PBMCs) should be isolated by density gradient centrifugation (e.g. Ficoll-Paque™, see "General Protocols" in the User Manuals or visit www.miltenyibiotec.com/protocols).

▲ **Note:** Remove platelets after density gradient separation: resuspend cell pellet in buffer and centrifuge at $200 \times g$ for 10–15 minutes at 20 °C. Carefully remove supernatant. Repeat washing step and carefully remove supernatant.

When working with tissues, prepare a single-cell suspension by a standard preparation method (see "General Protocols" in the User Manuals or visit www.miltenyibiotec.com/protocols).

▲ **Note:** Dead cells may bind non-specifically to MACS MicroBeads. To remove dead cells, we recommend using density gradient centrifugation or the Dead Cell Removal Kit (# 130-090-101).



2.2 Magnetic labeling

▲ Work fast, keep the cells cold, and use pre-cooled solutions. This will prevent capping of antibodies on the cell surface and non-specific cell labeling.

▲ Volumes for magnetic labeling given below are for up to 10^7 total cells. When working with fewer than 10^7 cells, use the same volumes as indicated. When working with higher cell numbers, scale up all reagent volumes and total volumes accordingly (e.g. for 2×10^7 total cells, use twice the volume of all indicated reagent volumes and total volumes).

▲ For optimal performance it is important to obtain a single-cell suspension before magnetic separation. Pass cells through 30 µm nylon mesh (Pre-Separation Filters # 130-041-407) to remove cell clumps which may clog the column.

▲ Primary PE-conjugated antibodies should be titrated to determine the optimal staining dilution. Staining should not increase fluorescence intensity of the negative population.

▲ The concentration of Anti-PE MultiSort MicroBeads used to achieve optimal magnetic separation and release of MultiSort MicroBeads is primarily dependent on the intensity of PE-conjugated antibody staining and, to some degree, also on the frequency of target cells in suspension. Dimly PE-stained target cells require a higher concentration of Anti-PE MultiSort MicroBeads to achieve optimal magnetic labeling and separation. Target cells with a high frequency (>50%) may also require a higher concentration of Anti-PE MultiSort MicroBeads than target cells with lower frequencies (see Table 2).

1. Determine cell number.
2. Centrifuge at $300 \times g$ for 10 minutes. Pipette off supernatant completely.
3. Resuspend cell pellet.
4. Stain cells with primary PE-conjugated antibody according to the manufacturer's recommendations. For MACS PE-conjugated antibodies, resuspend 10^7 total cells in 100 µL buffer and add 10 µL PE-conjugated antibody.
▲ **Note:** If second parameter sorting is performed indirectly, we recommend simultaneous labeling of the cells with the PE-conjugated primary antibody for the first parameter sorting and the e.g. APC-conjugated antibody for the second parameter sorting.
5. Mix well and incubate for 10 minutes in the dark at 4–8 °C or according to the manufacturer's recommendations.
▲ **Note:** Working on ice may require increased incubation times. Higher temperatures and/or longer incubation times lead to non-specific cell labeling.
6. Wash cells to remove unbound primary antibody by adding 1–2 mL of buffer per 10^7 cells and centrifuge at $300 \times g$ for 10 minutes.
7. (Optional) Repeat washing step.
8. Pipette off supernatant completely.
9. Resuspend cell pellet in buffer and add Anti-PE MultiSort MicroBeads according to recommendations in Table 2. For more details see also section 2.5.

Table 2: Buffer and MicroBead volumes per 10^7 total cells

Antigen Staining	Buffer	Anti-PE MultiSort MicroBeads
"dim"	80 µL	20 µL
"intermediate"	90 µL	10 µL
"bright"	96 µL	4 µL

▲ **Note:** A too strong dilution of the Anti-PE MultiSort MicroBeads may result in poor retention of the PE-labeled target cells. A too low dilution may result in less efficient release of MicroBeads.

10. Mix well and incubate for 15 minutes at 4–8 °C.
▲ **Note:** Working on ice may require increased incubation times. Higher temperatures and/or longer incubation times lead to non-specific cell labeling.
11. Wash cells carefully by adding 1–2 mL of buffer per 10^7 cells and centrifuge at $300 \times g$ for 10 minutes.
12. Pipette off supernatant completely.
13. Resuspend up to 10^8 cells in 500 µL of buffer.
▲ **Note:** For higher cell numbers, scale up buffer volume accordingly.
14. Proceed to magnetic separation (2.3).



2.3 Magnetic separation

▲ Choose an appropriate MACS Column and MACS Separator according to the number of total cells and the number of the magnetically labeled cells (see Table 1).

Magnetic separation with MS or LS Columns

1. Place column in the magnetic field of a suitable MACS Separator (see "Column data sheets").
 2. Prepare column by rinsing with appropriate amount of buffer:
MS: 500 μ L LS: 3 mL
 3. Apply cell suspension onto the column.
 4. Collect unlabeled cells which pass through and wash column with appropriate amount of buffer. Perform washing steps by adding buffer three times, each time once the column reservoir is empty.
MS: 3 \times 500 μ L LS: 3 \times 3 mL
Collect total effluent. This is the unlabeled cell fraction.
 5. Remove column from the separator and place it on a suitable collection tube.
 6. Pipette appropriate amount of buffer onto the column. Immediately flush out fraction with the magnetically labeled cells by firmly applying the plunger, supplied with the column.
MS: 1 mL LS: 5 mL
- ▲ **Note:** To increase the purity of the magnetically labeled fraction, it can be passed over a new, freshly prepared column.
7. Proceed to removal of MultiSort MicroBeads (2.4).

Magnetic separation with the autoMACS™ Separator

▲ Refer to the "autoMACS™ User Manual" for instructions on how to use the autoMACS Separator.

1. Prepare and prime autoMACS Separator.
2. Place tube containing the magnetically labeled cells in the autoMACS Separator. Choose a positive selection program according to the recommendations in the "autoMACS User Manual".
▲ **Note:** Program choice depends on the strength of magnetic labeling and the frequency of magnetically labeled cells. For details see autoMACS User Manual: "autoMACS Cell Separation Programs".
3. Proceed to removal of MultiSort MicroBeads (2.4).



2.4 Removal of MultiSort MicroBeads using MultiSort Release Reagent

1. Remove a sample to analyse the separation by flow cytometry and proceed with remaining magnetically labeled fraction.
2. Add 20 μ L MultiSort Release Reagent per 1 mL cell suspension
3. Mix well and incubate for 10 minutes at 4–8 °C.
4. (Optional) To remove any residual magnetically labeled cells, repeat the magnetic separation procedure as described in 2.3. Separate cells over a new column of the same type (MS or LS Columns) or use the same autoMACS program as will be used for the second separation. Collect magnetic (non-released) and non-magnetic (released) cell fractions to determine the efficiency of the release reaction (see 2.5).

▲ **Note:** This step is extremely important if the target cells of the second parameter sorting are rare after the first parameter sorting with the Anti-PE MultiSort MicroBeads (<10% target cells in the positive fraction after first sort).

5. Wash cells from the released fraction carefully by adding 1–2 mL of buffer per 10^7 cells and centrifuge at 300 \times g for 10 minutes.
6. Pipette off supernatant completely.
7. Resuspend cells in buffer in a final volume of 50 μ L per 10^7 total cells.
8. Add 30 μ L of MultiSort Stop Reagent per 10^7 total cells and mix well.
9. Add the recommended amount of direct or indirect MACS MicroBeads (see "MicroBead data sheets") to magnetically label the cells for the second marker. Adjust to 100 μ L total volume by adding buffer.
▲ **Note:** The Anti-PE antibody is of mouse IgG1 isotype. Thus, Goat Anti-Mouse IgG MicroBeads or Rat Anti-Mouse IgG1 MicroBeads cannot be used for second parameter sorting. When using other Anti-Immunoglobulin MicroBeads for the second parameter sorting, any reactivity with the isotype of the primary antibody of the first parameter sorting has to be excluded.
10. Mix well and incubate for 15 minutes at 4–8 °C.
11. Continue as described in the "MicroBead data sheet" and proceed to magnetic separation.

2.5 Recommendations to optimize cell isolation using Anti-PE MultiSort Kit

The efficiency of the release of the MultiSort MicroBeads is dependent on the strength of the magnetic labeling with Anti-PE MultiSort MicroBeads. The strength of the magnetic labeling depends on the intensity of staining with the primary PE-conjugated antibody and on the amount of Anti-PE MultiSort MicroBeads used for magnetic labeling (see Table 2). Typically, a release >90% is obtained when working with the Anti-PE MultiSort MicroBeads.

$$\text{Release (\%)} = \frac{100 \times (\text{No. of cells in released fraction})}{(\text{No. of cells in released} + \text{non-released fractions})}$$

Too strong magnetic labeling due to too high a concentration of the Anti-PE MultiSort MicroBeads results in an insufficient release of the Anti-PE MultiSort MicroBeads. An insufficient release decreases the purity of the cells after positive selection according to the second parameter.


- Use a lower amount of Anti-PE MultiSort MicroBeads if the recovery of the target cells after the primary parameter sorting with the Anti-PE MultiSort MicroBeads is sufficient but the release is <90%.

Too weak magnetic labeling due to too low a concentration of the Anti-PE MultiSort MicroBeads results in low recovery of the PE-labeled target cells in the positive fraction.

- Use more Anti-PE MultiSort MicroBeads if the release is efficient (>90%) but the recovery of the positive cells after the first positive selection with the Anti-PE MultiSort MicroBeads is too low.

3. Example of a separation using Anti-PE MultiSort Kit

Separation of human peripheral blood CD3⁺CD56⁺ cells using CD56-PE, Anti-PE MultiSort Kit and an MS Column for the first positive selection and CD3 MicroBeads and an MS Column for the second positive selection.



Refer to www.miltenyibiotec.com for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit www.miltenyibiotec.com/local to find your nearest Miltenyi Biotec contact.

Warnings

Reagents contain sodium azide. Under acidic conditions sodium azide yields hydrazoic acid, which is extremely toxic. Azide compounds should be diluted with running water before discarding. These precautions are recommended to avoid deposits in plumbing where explosive conditions may develop.

Legal notices

Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at www.miltenyibiotec.com, as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit www.miltenyibiotec.com for the most up-to-date information on Miltenyi Biotec products.

Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at www.miltenyibiotec.com for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

Trademarks

autoMACS, MACS, MidiMACS, the Miltenyi Biotec logo, MiniMACS, OctoMACS, QuadroMACS, SuperMACS, and VarioMACS are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide. All other trademarks mentioned in this publication are the property of their respective owners and are used for identification purposes only.

Ficoll-Paque is a trademark of GE Healthcare companies.

Copyright © 2020 Miltenyi Biotec and/or its affiliates. All rights reserved.